



**SMOBIO**

Small Bio, Smart Tool

[www.smobio.com](http://www.smobio.com)

## Product Information

ExcelTaq™ series

### 2X Fast Q-PCR Master Mix (SYBR, no ROX)

#### TQ1200 200 RXN

2X Fast Q-PCR Master Mix (SYBR, no ROX) 1 ml x 2

#### TQ1201 500 RXN

2X Fast Q-PCR Master Mix (SYBR, no ROX) 1 ml x 5

## Storage

**Aliquot to avoid multiple freeze-thaw cycles**

**Protect from light**

-20°C for 12 months

## Features

- High sensitivity and signal intensity
- Compatible with fast PCR program
- With smart blue contrast dye as a visual aid for reaction setup
- Low background
- High stability

## Description

The ExcelTaq™ 2X Fast Q-PCR Master Mix (SYBR, no ROX) is a ready-to-use reagent with all the essential components for quantitative real-time PCR (qPCR) except primers and templates. The master mix features high sensitivity and signal intensity (Fig. 1) as well as low background and better compatibility with fast PCR program (Fig. 2).

The ExcelTaq™ 2X Fast Q-PCR Master Mix (SYBR, no ROX) contains a highly stable hot-start *Taq* DNA polymerase in an optimized buffer with dsDNA specific SYBR green fluorescent dye. Consequently, this master mix features high stability during storage, even at 37°C for weeks.

With inert smart blue contrast dye, the ExcelTaq™ 2X Fast Q-PCR Master Mix (SYBR, no ROX) is ready-to-use and greatly reduces pipetting errors, while largely improving the reproducibility of the process. The ExcelTaq™ 2X Fast Q-PCR Master Mix (SYBR, no ROX) is also compatible with ROX reference dye if ROX is recommended by the manufacturer of the qPCR system. This master mix allows sensitive, precise amplification, real-time tracking of the amplification process, and simultaneous quantification for targeted DNA molecules.




Fig. 1. The amplification plot and standard curve of real-time PCR with cDNA template ranged from 25 ng to 1.4 fg in quantity, analyzed by using TQ1200 ExcelTaq™ 2X Fast Q-PCR Master Mix (SYBR, no ROX) for qPCR amplification.




Fig. 2. The overlapped amplification curves from different hot start duration display that TQ1200 ExcelTaq™ 2X Fast Q-PCR Master Mix (SYBR, no ROX) performs successfully in short duration of initial activation (2 min).

## Application

- Quantitative real-time PCR
- Quantitative two-step real-time PCR

## Instrument compatibility

- BioRad system:
  - CFX96
  - Chromo 4™ Real-Time Detector
  - DNA Engine Opticon™
  - DNA Engine Opticon™ 2
  - CFX384 Touch
- Cepheid system:
  - Smart Cycler®
- Eppendorf system:
  - Mastercycler® ep realplex
- Roche system:
  - Roche LightCycler® 480
  - Roche LightCycler® Nano
- QIAGEN system:
  - Rotor-Gene™ Q
- Illumina system:
  - Eco™

Note: ExcelTaq™ 2X Fast Q-PCR Master Mix (SYBR, no ROX) is compatible with a variety of real-time instruments, including but not limited to the list above.

## Recommended primer design

- Amplicon size: 80-150 bp
- T<sub>m</sub> value: around 60°C (calculated with Primer3 software)
- Primer length: 17-25 mer
- Sequence:
  - 45-55% of GC content is recommended.
  - Avoid regional high GC or AT content
  - Avoid palindrome sequence
  - Sequence with G or C at the 3' end is recommended.
- Specificity of primers should be confirmed through a BLAST search.

## Recommended reaction mixture set up for qPCR

Template	2 µl*
Forward primer	50 – 400 nM**
Reverse primer	50 – 400 nM**
2X Fast Q-PCR Master Mix (SYBR, no ROX)	10 µl
H <sub>2</sub> O	to 20 µl
<hr/>	
Total volume	20 µl

\*Final template concentration varies depending on the copy number of target present in the template solution. The recommended amount of template is: **100 fg -100 ng of cDNA, 80 pg -50 ng of gDNA or 10<sup>2</sup>-10<sup>8</sup> molecules of plasmid.**

\*\*The PCR primer concentration for an optimal qPCR reaction may vary according to primers' properties and template condition.

## Recommended qPCR program

Try Fast program first, and optimize the reaction conditions if necessary. If the Fast program still does not give optimal results, try the standard program.

### Fast program for qPCR

Steps	Temp.	Time	Cycles
Template denature and enzyme activation	95°C	20 sec	1
Denature	95°C	3 sec	40
Annealing /Extension	60°C	30 sec	
Melting curve analysis	Refer to instrument manual		

### Standard program for qPCR

Steps	Temp.	Time	Cycles
Template denature and enzyme activation	95°C	2 min	1
Denature	95°C	15 sec	40
Annealing/Extension	60°C	60 sec	
Melting curve analysis	Refer to instrument manual		

## **Other Information**

SMOBIO Technology, Inc. claims all warranties with respect to this document, expressed or implied, including but not limited to those of merchantability or fitness for a particular purpose. In no event shall SMOBIO Technology, Inc. be liable, whether in contract, tort, warranty, or under any statute or any other basis for special, incidental, indirect, punitive, multiple or consequential damages in connection with or arising from this document, including but not limited to the use thereof.

Caution: Not intended for human or animal diagnostic or therapeutic uses.

## Related Products

CK1000	Champion E. coli Transformation Kit
DM2300	ExcelBand 100 bp+3K DNA Ladder, 500 $\mu$ l
DM3100	ExcelBand 1 KB (0.25-10 kb) DNA Ladder, 500 $\mu$ l
DL5000	FluoroDye DNA Fluorescent Loading Dye (Green, 6 $\times$ ), 1 ml
NS1000	FluoroVue Nucleic Acid Gel Stain (10,000X), 500 $\mu$ l
PM2510	ExcelBand Enhanced 3-color Regular Range Protein Marker, 250 $\mu$ l $\times$ 2
RP1000	ExcelRT Reverse Transcriptase, 20,000 U
RP1100	ExcelRT One-step RT-PCR Kit, 50 RXN
RP1400	ExcelRT Reverse Transcription Kit II, 100 RXN
RI1000	RNAok RNase Inhibitor, 2000 U
TF1000	SMO-HiFi DNA Polymerase, 100 U $\times$ 1
TP5000	ExcelTaq Hot Start II DNA Polymerase, 500 U $\times$ 1
TP1200	ExcelTaq 5X PCR Master Dye Mix, 200 RXN
TQ1110	ExcelTaq 2X Q-PCR Master Mix (SYBR, ROX), 200 RXN
TQ2110	ExcelTaq 2X Q-PCR Master Mix (TaqMan, ROX), 200 RXN
WM1000	YesBlot Western Marker I, 250 $\mu$ l

For research use only

2019 ver. 1.3.1